

Product Certificate

Product:

| | |
|-------------------------|---|
| Product Code | MH-2150 |
| Product Description | Bone Marrow Collection System |
| Lot/Batch No. | 120904 |
| Method of Sterilization | Cobalt 60 Gamma, <u>SAL 10⁻⁶</u> |
| Date of Sterilization | 09/17/12 |
| Use Before (expiration) | September 2017 |

| TEST ITEM | Method | Specification | Conformance |
|----------------------------|-------------------|--|-------------|
| Product Appearance | Visual Inspection | Product contains proper components | Pass |
| Sterile Package Appearance | Visual Inspection | Product contains no damage punctures | Pass |
| Leak Test | Pressure Test | No leakage | Pass |
| Biological Test | Pyrogen Test | 20 EU/device limit – not to exceed limit | Pass |

Final Disposition: Acceptable

Approved by and Date:

 11/14/2012