

## Product Certificate

**Product:**

|                         |                                             |
|-------------------------|---------------------------------------------|
| Product Code            | MH-2150                                     |
| Product Description     | Bone Marrow Collection System               |
| Lot/Batch No.           | 151203                                      |
| Method of Sterilization | Cobalt 60 Gamma, <u>SAL 10<sup>-6</sup></u> |
| Date of Sterilization   | 1/18/16                                     |
| Use Before (expiration) | December 2020                               |

| TEST ITEM                  | Method            | Specification                            | Conformance |
|----------------------------|-------------------|------------------------------------------|-------------|
| Product Appearance         | Visual Inspection | Product contains proper components       | Pass        |
| Sterile Package Appearance | Visual Inspection | Product contains no damage punctures     | Pass        |
| Leak Test                  | Pressure Test     | No leakage                               | Pass        |
| Biological Test            | Pyrogen Test      | 20 EU/device limit – not to exceed limit | Pass        |

Final Disposition: Acceptable

Approved by and Date:

*Robert Capoluni 2/19/2016*