

Product Certificate

Product:

Product Code	MH-2150
Product Description	Bone Marrow Collection System
Lot/Batch No.	160909
Method of Sterilization	Cobalt 60 Gamma, SAL 10^{-6}
Date of Sterilization	9/19/2016
Use Before (expiration)	September 2021

TEST ITEM	Method	Specification	Conformance
Product Appearance	Visual Inspection	Product contains proper components	Pass
Sterile Package Appearance	Visual Inspection	Product contains no damage punctures	Pass
Leak Test	Pressure Test	No leakage	Pass
Biological Test	Pyrogen Test	20 EU/device limit – not to exceed limit	Pass

Final Disposition: Acceptable
Approved by and Date: <i>Robert J. Capolini</i> 10/10/2016